



Current Data Parameters
NAME 180A114_74
EXPNO 5
PROCNO 1

F2 - Acquisition Parameters
Date_ 20200317
Time 4.01 h
INSTRUM spect
PROBHD Z150364_0012 (
PULPROG hmbcetgp13nd
TD 2048
SOLVENT DMSO
NS 256
DS 16
SWH 5154.639 Hz
FIDRES 5.033827 Hz
AQ 0.1986560 sec
RG 186.12
DW 97.000 usec
DE 10.00 usec
TE 298.1 K
CNST6 120.0000000
CNST7 170.0000000
CNST13 8.0000000
D0 0.00000300 sec
D1 0.80000001 sec
D6 0.06250000 sec
D16 0.00020000 sec
IN0 0.00001810 sec
SPARSELI 0
TDav 1
SFO1 500.1124219 MHz
NUC1 1H
P1 12.00 usec
P2 24.00 usec
PLW1 13.00899982 W
SFO2 125.7659640 MHz
NUC2 13C
P3 10.00 usec
P24 2000.00 usec
PLW2 61.18099976 W
SPNAM[7] Crp60comp.4
SPOAL7 0.500
SPOFFS7 0 Hz
SPW7 9.34780025 W
GPNAM[1] SMSQ10.100
GPZ1 80.00 %
GPNAM[3] SMSQ10.100
GPZ3 14.00 %
GPNAM[4] SMSQ10.100
GPZ4 -8.00 %
GPNAM[5] SMSQ10.100
GPZ5 -4.00 %
GPNAM[6] SMSQ10.100
GPZ6 -2.00 %
P16 1000.00 usec

F1 - Acquisition parameters
TD 64
SFO1 125.766 MHz
FIDRES 863.259644 Hz
SW 219.649 ppm
FnMODE Echo-Antiecho

F2 - Processing parameters
SI 2048
SF 500.1100000 MHz
WDW SINE
SSB 4
LB 0 Hz
GB 0
PC 1.40

F1 - Processing parameters
SI 1024
MC2 echo-antiecho
SF 125.7527600 MHz
WDW QSINE
SSB 2
LB 0 Hz
GB 0

