



Current Data Parameters
NAME Feb26-2021-SY40
EXPNO 5
PROCNO 1

F2 - Acquisition Parameters

Date_ 20210227
Time 11.08 h
INSTRUM spect
PROBHD Z114607_0034 (
PULPROG hsqcedetgp
TD 1024
SOLVENT CDCl3
NS 36
DS 16
SWH 5952.381 Hz
FIDRES 11.625744 Hz
AQ 0.0860160 sec
RG 179.76
DW 84.000 usec
DE 6.50 usec
TE 292.6 K
CNST2 145.0000000
D0 0.00000300 sec
D1 1.50000000 sec
D4 0.00172414 sec
D11 0.03000000 sec
D13 0.00000400 sec
D16 0.00020000 sec
D21 0.00345000 sec
IN0 0.00001330 sec
TDav 1
ZGPTNS
SFO1 600.1323551 MHz
NUC1 1H
P1 11.00 usec
P2 22.00 usec
P28 1000.00 usec
PLW1 20.00000000 W
SFO2 150.9178988 MHz
NUC2 13C
CPDPRG[2] garp
P3 12.00 usec
P4 24.00 usec
PCPD2 60.00 usec
PLW2 106.00000000 W
PLW12 4.23999977 W
GPNAM[1] SMSQ10.100
GPZ1 80.00 %
GPNAM[2] SMSQ10.100
GPZ2 20.10 %
P16 1000.00 usec

F1 - Acquisition parameters
TD 256
SFO1 150.9179 MHz
FIDRES 293.703003 Hz
SW 249.102 ppm
FnMODE Echo-Antiecho

F2 - Processing parameters
SI 1024
SF 600.1300128 MHz
WDW QSINE
SSB 2
LB 0 Hz
GB 0
PC 1.40

F1 - Processing parameters
SI 1024
MC2 echo-antiecho
SF 150.9028085 MHz
WDW QSINE
SSB 2
LB 0 Hz
GB 0

