



Current Data Parameters
NAME FR-55
EXPNO 4
PROCNO 1

ppm F2 - Acquisition Parameters
Date_ 20200709
Time 3.00 h
INSTRUM spect
PROBHD Z44896_0022 (C
PULPROG hmbcgp1pndqf
TD 2048
SOLVENT CDCl3
NS 4
DS 16
20 SWH 7812.500 Hz
FIDRES 7.629395 Hz
AQ 0.1310720 sec
RG 186.75
40 DW 64.000 usec
DE 10.00 usec
TE 298.0 K
CNST2 145.0000000
CNST13 10.0000000
D0 0.00000300 sec
D1 1.50000000 sec
60 D2 0.00344828 sec
D6 0.05000000 sec
D16 0.00020000 sec
IN0 0.00001510 sec
80 TDav 1
SFO1 599.9687798 MHz
NUC1 1H
P1 12.00 usec
P2 24.00 usec
PLW1 3.59750009 W
100 SFO2 150.8764054 MHz
NUC2 13C
P3 15.00 usec
PLW2 90.00000000 W
GPNAM[1] SMSQ10.100
120 GPZ1 50.00 %
GPNAM[2] SMSQ10.100
GPZ2 30.00 %
GPNAM[3] SMSQ10.100
GPZ3 40.10 %
P16 1000.00 usec
140 F1 - Acquisition parameters
TD 128
SFO1 150.8764 MHz
FIDRES 517.384094 Hz
SW 219.468 ppm
FnMODE QF
160 F2 - Processing parameters
SI 2048
SF 599.9650000 MHz
WDW SINE
SSB 0
180 LB 0 Hz
GB 0
PC 1.40
200 F1 - Processing parameters
SI 1024
MC2 QF
SF 150.8613192 MHz
WDW SINE
SSB 0
LB 0 Hz
GB 0

